



Indian Institute of Technology Hyderabad

Kandi- 502 284, Sangareddy, Telangana State, India

Phone: 040 2301 6823: Fax: 040 2301 6032

Email: office.stores@iith.ac.in

GST Number: 36AAAI2661J1ZP

Date: 09/06/2022

CORRIGENDUM /ADDENDUM – 01

1. It is notified to all concerned parties that with reference to our Tender No IITH/JICA/5(5.14)/BUMA/CE/2022/0/T011 Dated 26.5.2022 the following points are clarified, and they may be considered as Corrigendum /Amendments to the specifications given in the document -

I. SYSTEM TECHNICAL SPECIFICATIONS OF PRIMARY SYSTEM

- I.1. The stroke of load frame stroke is modified to a minimum of 70mm.
I.4. The operating temperatures of submersible load cells is modified to +5°C to +50°C
I.5. The displacement range mentioned is modified to a range of 25mm
I.6. Pressure sensor accuracy is modified to +/- 0.25% of Full Range Output.

II. OPTIONAL FEATURES TO BE INCLUDED WITH THE PRIMARY SYSTEM

Include one displacement transducer 50mm range for measurements on large size triaxial specimens.

II.A. TECHNICAL SPECIFICATIONS OF UNSATURATED SOIL TESTING

Unsaturated triaxial setup is for sample size of 50mm diameter x 100mm height and in static testing only

II.A. Technical specifications of unsaturated soil testing, it is clarified that unsaturated triaxial set up will consist of

- Double wall (DW) triaxial cell for Unsaturated Triaxial Testing with inner acrylic and suitable for use with internal submersible load cells. It will have sufficient exit ports and connections (ports) for internal cables and for air/water flow
- Volume change unit to measure the inner chamber pressure
- Submersible load cell of 10 kN capacity specific to DW cell.

Bid Submission Close Date

Extended up to 20/06/2022 by 14:00 hrs

2. All other terms and conditions of the tender remain unchanged. Bidders, who have already submitted their bids prior to issue of this corrigendum need to submit again.