

Date: 14 Jun 2018


**CORRIGENDUM – 1**

1. It is notified to all concerned parties that with reference to our Tender No. IITH/5(1)/AKR/2018/T009 dated 28.05.2018 for supply of "MICRO CT IMAGING SYSTEM", the following changes are being made w.r.t. the Tender: -

SL. NO.		EXISTING	AMENDED AS
(i)	Page no. 10 of 14 of the tender document – Annexure I	<p>Camera Specifications:</p> <p><b><u>S. no. 6</u></b> CCD Camera with field view of 68mm or better</p> <p><b><u>S. no. 7</u></b> The system should be equipped with a very high resolution CCD camera with minimum 11 mega-pixels.</p>	<p>Camera Specifications:</p> <p><b><u>S. no. 6</u></b> High Resolution CCD/CMOS Flat Panel Camera with field view of 68mm or better to provide high degree of sensitivity and resolution.</p> <p><b><u>S. no. 7</u></b> In case CCD Camera is offered, the camera should be with 11 mega-pixels or higher resolution.</p>
<b>Revised Due Date &amp; Time for Submission of Tenders As below:-</b>			
(ii)	<b>Last Date &amp; Time for submission of Tender</b>	18 Jun 2018 by 03.00 pm	28 Jun 2018 by 03.00 pm
	<b>Date &amp; Time of Tender Opening</b>	18 Jun 2018 by 03.30 pm	28 Jun 2018 by 03.30 pm

2. All other terms and conditions of the tender remain unchanged. Bidders, who have already submitted their bids prior to issue of this corrigendum, may opt to quote afresh also. In case more than one bid is received from one firm, the bid which is received latest, before the extended Due Date shall be considered for all purposes.

Yours faithfully,



(V VENKAT RAO)  
Joint Registrar  
IIT Hyderabad