

Date: 16 Feb 2018

**CORRIGENDUM – 1**


1. It is notified to all concerned parties that with reference to our **Tender No. IITH/5(11)/JMOHANTY/2017/T191 dated 30.01.2018 for supply of "MULTIMODAL SCANNING PROBE MICROSCOPE"**., the following changes are being made w.r.t. the Tender: -

S.No.	Description	EXISTING	AMENDED AS
(i)	Page No. 10 of 14 of the Tender Document	<p><b><u>Point Nos. 2, 3 &amp; 4</u></b>            2. XY scan range should be more than 90µmx90µm. Smaller scanner size can be quoted as option.            3. Z scanner should have range more than 15µm.            4. It must accommodate sample size of 50x50x20mm. For smaller samples vacuum chunk must be provided..</p>	<p><b><u>Point Nos. 2, 3 &amp; 4</u></b>            2. XY scan range should be 90µmx90µm or more. Smaller scanner size can be quoted as option.            3. Z scanner should have range more than 10µm or more.            4. It must accommodate sample size of 50x50x15mm or more. For smaller samples vacuum chunk must be provided.</p>
(ii)	<b>Revised Due Date &amp; Time for Submission of Tenders is as below: -</b>		
	Last Date & Time for submission of Tender	22 Feb 2018 by 03.00 pm	02 Mar 2018 by 03.00 pm
	Date & Time of Tender Opening	22 Feb 2018 by 03.30 pm	02 Mar 2018 by 03.30 pm

2. If any of you have already submitted your bid prior to issue of this corrigendum you are required to re-submit your bid before the revised Due Date & Time, considering changes in the specifications of the Tender as per this Corrigendum.

3. All other Terms & Conditions of Tender remain completely unchanged.

Yours faithfully,

  
(V.VENKAT RAO)  
Joint Registrar  
IIT Hyderabad